

INSPECTION PROTOCOL

Variotec® 460 Tracergas



Serial no. (e.g.: 065 15 00480)

--	--	--

02.05.2012

1.0	General status																		
1.1	– Perfect condition (e.g.: Y / N)																		
1.2	– Fine dust filter correct (e.g.: Y / N)																		
1.3	– Disposable/rechargeable battery capacity (e.g.: ¼)																		

2.0	Pump check																		
2.1	– Pump error F100 in seal																		

3.0	Gas-sensitive semiconductor (Inspection above ground/House)																		
3.1	Zero point – Display with fresh air																		
3.2	Test gas 100 ppm H ₂ (in synthetic air) – Display 70 – 150 ppm																		

4.0	Thermal conductivity sensor (Measuring in bar holes/Gas measuring)																		
4.1	Zero point – Display -1.0 – +1.0 % vol.																		
4.2	Test gas 5 % vol. H ₂ (in N ₂) – Display 3.0 – 7.0 % vol.																		

5.0	Rapid test with PGG H2 (Inspection above ground/House)																		
5.1	Zero point – Display with fresh air																		
5.2	Test gas H2 (from PGG H2) – Display >1.5 ppm; 8.0 ppm typical																		

6.0	Comments – Housing damaged – Adjustment, repair – Inspection at factory – or similar																		
------------	---	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--

7.0	Inspection																		
	– Day																		
	– Month																		
	– Year																		
	– Signature																		

