

Inspection protocol

Sensor (e.g. CO):

test gas conc. (e.g. 40 ppm CO):

display tolerance zero point (e.g. -3 – +3 ppm)

display tolerance test gas (e.g. 27 – 33 ppm)

Serial no. (e.g. 062 01 0001):

EX-TEC® GM 4 / OD 4



(Display tolerances according to
chapter 5.4 of operating instruction)

22.07.2008

1.0	Device status										
1.1	- status correct (e.g.: Y / N)										
1.2	- remaining operating hours (e.g.: 5 h)										

2.0	Pump test										
2.1	- fault F 100 when sealing										
2.2	- pump filter changed (e.g.: Y / N)										

3.0	Display accuracy										
3.1	- display (zero point)										
	- within tolerance (e.g.: Y / N)										
3.2	- display (test gas)										
	- within tolerance (e.g.: Y / N)										

4.0	AL 1 alarm triggering										
4.1	optical alarm (e.g.: Y / N)										
4.2	audible alarm (e.g.: Y / N)										

5.0	Observations										
	- housing broken										
	- adjustment, repair										
	- factory inspection										
	- or the like										

6.0	Test										
	- day										
	- month										
	- year										
	- signature										